

Certificate of Analysis



Client:
Peptide Partners

Laboratory:
TrustPointe Analytics LLC
1743 142nd Ave Suite 4
Dorr, MI 49323



Verify at TrustPointeLIMS.com
Key=FA47EYEW9WVM

Sample Information

Sample Name	SS-31 50mg, 1 vial, ENDO	Client SID	SS2611	Published	2026-08-14 15:11
--------------------	--------------------------	-------------------	--------	------------------	------------------

Results

Sample IDs	SPL-4633-R01
USP <85> Bacterial Endotoxin	10.1 EU/mL
USP <85> Sample CV %	2.4 %
USP <85> Spike CV %	3.0 %
USP <85> Spike Recovery	87 %

Comments for SPL-4633-R01

General

Original dilution 2mL; USP limit 5 EU/kg/hr

Approvals

Created by

Ashlee Lust
President

Reviewed by

Joshua Lust
Managing Director