KOVERA LABS

BACTERIAL ENDOTOXIN ANALYSIS

(Kinetic Chromogenic LAL, USP <85>)

Sample ID: **KVR-2026-18A622**

CLIENT & SAMPLE INFORMATION

Client	Peptide Partners	**Analysis Date**	04/14/2026
Product Name	Cagrilintide	**Strength**	10 mg
Lot / Batch	CG202603	**Condition**	Lyophilized

This Certificate of Analysis certifies that the sample listed herein was tested for bacterial endotoxin using a kinetic chromogenic LAL assay in accordance with USP <85> Bacterial Endotoxins Test.

TEST METHODOLOGY

Test Performed	Quantitative Bacterial Endotoxin Test	**Compendial Ref**	USP <85>
Assay Type	Chromogenic (Kinetic)	**Detection λ**	405 nm
Detection Range	0.01 – 1.0 EU/mL	**Endotoxin Std**	E. coli O111:B4
Dilution Vol	2.0 mL	**Matrix**	LAL Reagent Water

QUANTITATIVE RESULTS

Parameter	Result
Endotoxin Level	**< 0.06 EU/mL**
Acceptance Limit	**≤ 0.5 EU/mL**
Sample CV (%)	**2.6%**
Spike CV (%)	**4.8%**
Spike Recovery (%)	**109%**
Final Determination	**PASS**

CONTROLS

Control	Expected	Observed	Status
Positive Control	Detectable	As expected	**Pass**
Negative Control	No signal	As expected	**Pass**

INTERPRETATION

Endotoxin content was quantitatively determined using a kinetic chromogenic LAL assay in accordance with USP <85>. Result reported as below quantitation threshold. The measured endotoxin level meets the specified acceptance criteria. The sample passes the endotoxin limit test.

AUTHORIZATION

PREPARED BY

Liam Clarke, B.S.

QC Analyst

REVIEWED BY

David Cohen, M.S.

Quality Director